

**EC Certificate**  
**Directive 93/42/EEC Annex V**  
**Production Quality Assurance**  
**Medical Devices**

**Registration No.:** DD 60113931 0001

**Report No.:** 17032312 005

**Manufacturer:** Shenzhen Calibeur Industries  
Co., Ltd.  
No. 11 Bldg., Shiguan Industrial Park  
Gongming, Guangming New District  
Shenzhen  
518107 Guangdong  
China

**Products:** Infrared Thermometers  
Replaces Approval, Registration No.: DD 60088530 0001

**Expiry Date:** 2018-09-24

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

**Effective Date:** 2016-12-02

**Date:** 2016-12-02



**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.